

Instruction for Use KL03/IFU-C1

Surgicare

01 Product Description

Product Name: Powder Free Polymer Coated Sterile Latex

Surgical Gloves

Material : Natural Rubber Latex Colour : White to Pale Yellow

Shape : Anatomic Cuff : Beaded

External Surface: Antitack Polymer treated

Internal Surface: Polymer Coated

Size : 5.5, 6.0, 6.5, 7.0, 7.5, 8.0, 8.5 and 9.0 Sterilization : Ethylene Oxide (EO) / GAMMA (R)

Shelf Life : 5 years

Basic UDI-DI : 806363LSGFM4

02 Intended Use

This disposal medical device is made up of natural rubber latex which is anatomically shaped with thumb position towards the Palm side of the index finger which reduces the fatigue on the hands, intended to be worn on the hands usually in surgical settings to provide barrier against potentially infectious fluids and other contaminants.

03 Product Classification & Standard compliance

Medical Device Classification: Class IIa, Rule#06

Conformity assessment route: Annex II section 4 of council

directive 93/42/EEC

Regulatory Authority : DNV Product Assurance As

Notified Body Number : 2460

Standard Compliance :

EN 455-1:2020, EN 455-2:2015, EN 455-3:2015, EN 455-4:2009, EN ISO 15223-1:2021, EN 1041:2013, EN 566-1:2001, EN ISO 11135: 2014/A1:2019, EN ISO 11137-1:2015/A2:2019, EN ISO 11737-1:2018/A1:2021, EN ISO 11737-2:2020, ISO 11607-1:2019, ISO 11607-2:2019, EN ISO 10993-1:2020, EN ISO 10993-5:2009, EN ISO 10993-7:2008, EN ISO 10993-10:2021, EN ISO 10993-11:2017, EN ISO 10993-23:2021, ISO 10282:2014, ASTM D 3577:2019, ISO 13485:2016, ISO 14001:2015, ISO 9001:2015.

04 Storage Instruction

Gloves must be stored in cool dry environment which is dust free. Cartons and Boxes must be stored unopened until required. Recommended storage temperature is 5°C-35°C. Avoid exposure to direct light, heat and excessive humidity.

As ozone is deleterious, storage area should not contain any equipment capable of generating ozone such as ultraviolet light, fluorescent lights, mercury vapour lamps, photocopier, high voltage equipment, x-ray units, electric motors and electro surgical equipments.

05 Indication For Use

- Dry hands thoroughly before donning.
- Protective gloves should only be used for the intended application and in the correct size.
- These are sterile gloves for single use only.
- Users should take care when using the gloves.
 Using them solely according to their intended application.
- Before usage, inspect the gloves for any defect or imperfection.

06 Contraindication

- Tatex gloves are made of Natural rubber latex, which may cause allergic reactions including anaphylaxis response if the user is allergic to latex.
- Gloves contain Natural Latex; persons who are sensitive to Latex should consult a physician before using.



Instruction for Use KL03/IFU-C1

Surgicare

07 Precautions


- Do not use package is damaged or wet.
- Risk of reuse: May cause infection, allergic reaction and poor barrier protection.
- Gloves shall not be worn where there is a risk of entanglement by moving parts of machines is needed.
- The results do not reflect the actual duration of protection in the workplace due to other factors influencing the performance, such as temperature, abrasion, degradation etc.

08 Warnings

- Device disposal should be done as per local regulatory norms.
- Do not re-sterilize.
- The product contains Natural Rubber Latex which may cause allergic reactions including anaphylactic responses to some individuals.
- The gloves not intend to prevent Electrical shock care should be taken to have proper earthing in Medical Device Electrical appliances user.
- Necessary caution to be practiced against probable Electrical Hazards.

09 Directions for use

Glove Opening and Donning Procedure:



- Remove the walleted gloves (inner wrapper) from the Pouch (outer wrapper) by Peel open from the corner for Paper Pouch (Peel down to open pouch).
- Open the walleted glove to see "Left" and "Right" compartment.
- Pinch back upper and lower flaps of the inner wrapper.
- Using the middle flaps, open the wrapper touching only the 1 inch margin for safety.
- Be sure wrapper does not close over gloves after opening to avoid contamination.
- Using the thumb and the first two fingers of the nondominant hand, pinch the cuff of the folded edge of the glove cuff for the dominant hand, touching only the inside surface of the glove.
- Slide dominant hand in to the gloves keeping hand point downwards and pull up to wrist.
- Using the glove hand insert the 4 fingers under the cuff of the other glove and pull the glove up to the arm.
- Adjust the gloves as necessary.

Glove Removal Procedure:



- Take hold of the first glove at the wrist.
- Fold it over and peel it back, turning it inside out as it goes. Once the glove is off, hold it with your gloved hand.
- To remove the other glove, place your bare fingers inside the cuff without touching the glove exterior. Peel the glove off from the inside, turning it inside out as it goes. Use it to envelope the other glove.



Instruction for Use | Surgicare KL03/IFU-C1

Personal Protective Equipment

EN ISO 374-1/Type A EN ISO 374-1/Type B EN ISO 374-1/Type C









TYPE A Permeation min. 6 chemicals, level 2 (table 1) TYPE B Permeation min. 3 chemicals, level 2 (table 1)

TYPE C Permeation min. 1 chemical, level 1 (table 1)



Protective gloves protecting from viruses, bacteria and fungus



Protective gloves protecting from bacteria and fungus

Performance Level

T CHOTHLANCE LEVEL			
	Time	Protection index	
	>10 min	1	
	>30 min	2	
	>60 min	3	
	>120 min	4	
	>240 min	5	
	>480 min	6	

Table 1

Chemicals

- Methanol В
- Acetone
- \mathbf{C} Acetonitrile
- D Dichloromethane
- Е Carbon disulphide
- F Toluene
- G Dimethylamine
- Η Tetrahydrofuran
- Ethyl acetate
- n-Heptane
- Sodium hydroxide 40% K
- Sulphuric acid 96%
- Nitric acid 65% Μ
- Acetic acid 99%
- O Ammonium hydroxide 25%
- Hydrogen peroxide 30%
- S Hydrofluoric acid 40%
- Formaldehyde 37%

EN ISO 374-1/Type C













In Compliance with the Harmonized Standards Personal Protective Equipment, risk category III under Regulation (EU)2016/425

EN ISO 21420: 2020 EN ISO 374-1: 2016

EN 374-2:2014 EN 16523-1:2015

EN 374-4: 2013 EN ISO 374-5:2016

Regulatory Authority : SGS Fimko Oy

Notified Body Number: 0598

Sodium hydroxide 40%	Level 6
Hydrogen peroxide 30%	Level 6
Formaldehyde 37%	Level 1

11 Explanation of Symbols



Manufacturer

EC REP

Authorized representative in the European Community AMSTERMED B.V.,

Saturnusstraat 46-62,

Unit 032, 2132 HB Hoofddorp,

The Netherlands

Date of Manufacture / Country of Manufacture

Use by date

LOT

Lot No

REF

Reference Number / Catalogue Number

SN

Serial Number

 ϵ

CE Logo

STERILE Sterilization using Ethylene oxide

STERILE R Sterilization using Irradiation



Single Sterile Barrier System

Do not re-sterilize



Do not use package is damaged or wet



Keep away from sunlight



Keep dry



Temperature limit



This way up or end up



Keep away from Ozone



Single Use



Instruction for Use



Caution



Latex Caution



Medical Device



Unique Device Identifier