

Product : Powder Free Non – Sterile Synthetic Nitrile Examination Gloves

Product Description

Material : Synthetic Nitrile Rubber Latex
 Colour : Blue
 Shape : Ambidextrous
 Cuff : Beaded
 External Surface: Textured and Chlorinated
 Internal Surface: Polymer coated
 Shelf Life : 3 Years
 Basic UDI DI : 806363SNGNFF3

Product Classification

Medical Device : Class I, Rule 05

Product Compliance

- ✓ EN 455 Parts 1,2,3 & 4
- ✓ ASTM D 6319, IS 15354, ISO 11193-1
- ✓ EN ISO 15223-1, EN 1041
- ✓ EN ISO 10993 Parts 1,5,10 & 23
- ✓ ISO 9001, ISO 14001, ISO 13485, EN ISO 14971

Product Packaging

➔ 10 boxes x 100 pcs / 1000 pcs per Carton

Product Safety

Freedom from Holes : Inspection Level 1, AQL 1.5
 Visual Major Defects : Inspection Level 1, AQL 1.5
 Visual Minor Defects : Inspection Level1, AQL 2.5
 Powder Amount : < 2 mg/glove



Physical Dimension

Size	Minimum Length (mm)	Palm Width (mm)	Weight/ glove (g)
X Small	290	70 ± 10	4.5 – 5.0
Small	290	80 ± 10	5.0 – 5.5
Medium	290	95 ± 10	5.5 – 6.0
Large	290	110 ± 10	6.0 – 6.5
X-Large	290	120 ± 10	6.5 – 7.0

Thickness

Location	Minimum Thickness (mm)	
	ASTM D 6319	Kanam
Finger (13±3 mm from finger tip)	0.05	0.14
Palm (Center of the palm)	0.05	0.12

Physical Properties

Characteristics	Before Ageing			After Ageing (70±2°C for 166±2 h)		
	ASTM D 6319	EN 455-2	ISO 11193-1 & IS 15354	ASTM D 6319	EN 455-2	ISO 11193-1 & IS 15354
Tensile Strength (Mpa)	14 min	-	-	14 min	-	-
Ultimate Elongation (%)	500 min	-	500 min	400 min	-	500 min
Stress @ 500% Elongation (Mpa)	-	-	-	N/A	-	N/A
Force @ Break (N)	-	6 min	7 min	-	6 min	7 min