

# EU TYPE-EXAMINATION CERTIFICATE

## Regulation (EU) 2016/425, MODULE B

### 0598/PPE/22/3430 Issue 2

<b>Product</b>	Protective gloves against chemicals and micro-organisms
<b>Model</b>	Powder free latex surgical gloves as an underglove
<b>Trademark</b>	Surgicare® Sensitive / Surgicare® Underglove
<b>Certificate Holder / Manufacturer</b>	Kanam Latex Industries Private Limited 12/67 C, Ananthanadarkudy, Asaripallam P.O., Nagercoil – 629 201, Kanyakumari District, Tamilnadu, India
<b>Standard(s)</b>	Product complies with the applicable essential health and safety requirements of Regulation (EU) 2016/425 and standard(s) mentioned below EN ISO 374-1:2016+A1:2018 (Protective gloves against chemicals and micro-organisms) Type C glove against reagents: K, L, T; EN ISO 374-5:2016 (Protective gloves against chemicals and micro-organisms) gloves protecting against bacteria and fungi (EN 374-2:2019 inspection level 2 (AQL <1.5 G1), and virus (ISO 16604)); EN ISO 21420:2020 (Protective gloves - General Requirements).
<b>Other Information</b>	This certificate shall be used in conjunction with conformity assessment procedure module C2 or D.
<b>Validity</b>	This certificate is valid until 2028-08-03.
<b>Date of issue</b>	2023-08-03
<b>Signature</b>	<b>SGS Fimko Ltd</b>



Mikko Hirvonen  
Senior Specialist

SGS Fimko Ltd is a Notified Body (0598) according to the Personal Protective Equipment Regulation (EU)



**Additional information** Powder free latex surgical gloves as an underglove, anatomically shaped natural rubber latex glove, with beaded cuff, single use, which are available in seven nominal hand sizes: 6, 6½, 7, 7½, 8, 8½ & 9.

Sold under brand names: Surgicare® SENSITIVE & Surgicare® UNDERGLOVE.

**Contract No:** INP-CERT230700083-01 / GB-202306004028.

**Technical File:** TF/PPE/LUG-PF, Issue 02, Rev 09, Date: 22.06.2023.