

EU TYPE-EXAMINATION CERTIFICATE

Regulation (EU) 2016/425, MODULE B

0598/PPE/21/2869 Issue 1

Product Protective gloves against chemicals

Model Powdered Sterile Latex Surgical Gloves for Gynaecological Use (400 mm & 480 mm)

Trademark Surgicare

Certificate Holder / Manufacturer Kanam Latex Industries Private Limited
12/67 C, Ananthanadarkudy, Asaripallam P.O., Nagercoil – 629 201, Kanyakumari District, Tamil Nadu, India

Product complies with the applicable essential health and safety requirements of Regulation (EU) 2016/425 and standard(s) mentioned below

Standard(s) EN ISO 374-1:2016 (Protective gloves against chemicals and micro-organisms) as a Type C glove against reagents: K, P & T;
EN ISO 374-5:2016 (Protective gloves against chemicals and micro-organisms) gloves protecting against bacteria & fungi (EN 374-2:2014, inspection level 2 (AQL <1.5 G1));
EN 420:2003/A1:2009 (Protective gloves general requirements).

Other Information

This certificate shall be used in conjunction with conformity assessment procedure module C2 or D.

Validity This certificate is valid until 2026-12-20.

Date of issue 2021-12-20

SGS Fimko Ltd

Signature



SGS Fimko Ltd is a Notified Body (0598) according to the Personal Protective Equipment Regulation (EU)



Additional information Surgicare® (logo) Gynaecological Gloves, Powdered latex surgical gloves for gynaecological use, creamy white natural rubber latex gloves, powdered internal surface, textured external surface, beaded cuff, sterile, available with nominal lengths of 400mm (shape: anatomic) or 480mm (shape: hand specific) and in three nominal hand sizes: 6.5 (S), 7.5 (M), & 8.5 (L).

Contract No: GB-CERT210900201 / GB-202106000145 / 238716.

Technical File: TCF/PPE/LGG-PP, Issue 01, Rev 03, dated 18.08.2021.
(Previously certified under certificate number: GB19/964090).