

# EU TYPE-EXAMINATION CERTIFICATE

## Regulation (EU) 2016/425, MODULE B

### 0598/PPE/23/4061 Issue 2

**Product** Protective gloves against chemicals and micro-organisms

**Model** Powdered Latex Surgical Gloves (Sterile & Non Sterile)

**Trademark** -

**Certificate Holder / Manufacturer** Kanam Latex Industries Private Limited  
12/67 C, Ananthanadarkudy, Asaripallam P.O, Nagercoil – 629 201, Kanyakumari District, Tamil Nadu, India

**Standard(s)** Product complies with the applicable essential health and safety requirements of Regulation (EU) 2016/425 and standard(s) mentioned below

EN ISO 374-1:2016+A1:2018 (Protective gloves against chemicals and micro-organisms) Type B glove against reagents: K, L, M, T;

EN ISO 374-5:2016 (Protective gloves against chemicals and micro-organisms) gloves protecting against bacteria and fungi (EN 374-2:2019 inspection level 3 (AQL <0.65 G1), and virus (ISO 16604:2004));

EN ISO 21420:2020 (Protective gloves - General Requirements).


**Other Information** This certificate shall be used in conjunction with conformity assessment procedure module C2 or D.

**Validity** This certificate is valid until 2029-06-25.

**Date of issue** 2024-06-25

**SGS Fimko Ltd**

**Signature**



Kelvin Shepherd

SGS Fimko Ltd is a Notified Body (0598) according to the Personal Protective Equipment Regulation (EU)



**Additional information** KANAM LATEX INDUSTRIES PRIVATE LIMITED, Powdered Latex Surgical Gloves (sterile & non-sterile), Creamy white natural rubber latex glove, anatomically shaped, textured, powdered, beaded cuff, available in seven nominal hand sizes: 6.0, 6.5, 7.0, 7.5, 8.0, 8.5 & 9.0.

Sold under brand names: Surgicare, Surgicare E, Kaltex, Kaltex EC, Surgicare Lowpro, Mumu, Dona Sensitive, Med-Comfort, Comfort, Comfort Powdered.

**Contract No:** INP-CERT240300036-01/GB-202403005443.

**Technical File:** TCF/PPE/LSG-PP, Issue 02, Rev 09, Dated: 30.05.2024.