

EU TYPE-EXAMINATION CERTIFICATE

Regulation (EU) 2016/425, MODULE B

0598/PPE/25/2766 Issue 1

Product Protective gloves against dangerous chemicals and micro-organisms

Model Powder Free Latex Micro Surgical Gloves (Sterile & Non-Sterile)

Trademark Surgicare Micro, Surgicare Neuro, Dermagel Micro, Microtex

Certificate Holder / Manufacturer Kanam Latex Industries Private Limited
12/67 C, Ananthanadarkudy, Asaripallam P.O, Nagercoil – 629 201, Kanyakumari District, Tamil Nadu, India

Product complies with the applicable essential health and safety requirements of Regulation (EU) 2016/425 and standard(s) mentioned below

Standard(s) EN ISO 374-1:2016+A1:2018 (Protective gloves against dangerous chemicals and micro-organisms) Type B glove against reagents: K, M, T;
EN ISO 374-5:2016 (Protective gloves against dangerous chemicals and micro-organisms) gloves protecting against bacteria and fungi (EN 374-2:2019 inspection level 3 (AQL <0.65 G1), and virus (ISO 16604));
EN ISO 21420:2020 (Protective gloves - General Requirements).


Other Information This certificate shall be used in conjunction with conformity assessment procedure module C2 or D.

Validity This certificate is valid until 2030-04-24.

Date of issue 2025-04-24

SGS Fimko Oy

Signature



Kelvin Shepherd

SGS Fimko Oy is a Notified Body (0598) according to the Personal Protective Equipment Regulation (EU)



Additional information Powder Free Latex Micro Surgical Gloves (Sterile & Non-Sterile), light brown natural rubber latex glove, anatomically shaped, beaded cuff, textured external surface, polymer coated internal surface, single use, available in eight nominal hand sizes: 5.5, 6.0, 6.5, 7.0, 7.5, 8.0, 8.5 & 9.0.

Sold under brand names: Surgicare Micro, Surgicare Neuro, Dermagel Micro, Microtex.

Contract No: INP-CERT250100005-01/GB-202410007357.

Technical File: TCF/PPE/LSG-MICRO, Rev 01, Dated: 24.03.2025.